

Clinical Commissioning of Eclipse Treatment Planning System for Protons at PTCH

Narayan Sahoo PhD, X. Ronald Zhu, PhD, MingFwu Lii, MS, Xiaodong Zhang, PhD, Richard Amos, MS, Richard Wu, MS, John R. Zullo, BS, Matthew Barfield, Bijan Arjomandy, PhD, Uwe Titt, PhD, Wayne Newhauser, PhD, Radhe Mohan, PhD and Michael Gillin, PhD

Dept. of Radiation Physics, UTMDACC

Objective

- Review our experience in collecting the required proton beam input data for Eclipse Treatment Planning System
- Compare the measured versus calculated dose distributions in a simple water phantom

Methods and Materials

- Eclipse treatment planning system requires a limited number of both in-phantom and in-air measurements for passively modulated proton beam
- Eclipse uses a Gaussian pencil beam model for dose calculation
- The knowledge of depth dose curve and proton fluence is needed for calculation of dose in patient

Measurements in water

Pristine Bragg Peaks (PBP)-PDD of unmodulated beam

Methods and Materials

Measurements in air

Z-fluence - open field longitudinal profiles

Effective source position is derived by fitting the fluence to inverse square law

Open field transverse profiles to determine field sizes at different distances

Virtual source position is derived from measured beam divergence

Half-blocked field transverse profiles

The effective size of the proton source is calculated using the penumbra width of the half-blocked field


Measurements

Eclipse requires measurements at the thinnest and a thick section of the stationary RMW


PBP and Z-fluence

- PTW MP3 Scanning system
- Advanced Markus Chamber
- Point by point integration of charge for 4 seconds
- Minimum five points of measurement for Z-fluence including isocenter
- Used fixed MU for Z-fluence measurements for fixed gantry room

Pristine Bragg Peaks, 250 MeV Proton Beam at PTCH



Z_fluence for different enrgies



In-air transverse profiles


- Kodak EDR2 films were used to measure the cross beam profiles in air
- The film was read using a Vidar scanner and was analyzed with Scanditronix OnmiPro software
- Good agreement between pin-point chamber and film measurement validated the accuracy of the film dosimetry

Film vs Ion Chamber



SSD, Penumbra and width are in cm

FILM	MU	SSD	Dose	Pen. Left	Pen. Right	Width 50%
	200	288	172.3	1.31	1.17	19.09
	200	270	198.0	0.97	0.88	17.97
	200	250	229.3	0.55	0.5	16.69
				Penumbra 80%-20%		
ION CHAMBER	MU	SSD	Dose	Left	Right	Width 50%
		288		1.24	1.19	19.04
		270		0.97	0.92	17.92
		250		0.59	0.57	16.63
				Penumbra 80%-20%		
DEVIATIONS	MU	SSD	Dose	Left	Right	Width 50%
		288		0.07	-0.02	0.05
		270		0.00	-0.04	0.05
		250		-0.04	-0.07	0.06


Open beam cross plane in-air profiles



Symbols = measurements
Solid lines = Monte Carlo



Half-beam cross plane in-air profiles for 250 MeV beam




Materials and Methods


- A total of ten measurements are required for each option. These ten measurements could be completed in approximately ten hours.
- Measured data were converted into the required W2CAD format and were entered into the Eclipse Treatment Planning system.
- After a complete set of data was entered for each option, dose calculations were performed in a water phantom to test the accuracy of the beam modeling in Eclipse

Results-SOBP


G1, RMW08, 250 MeV, SOBP 10 cm




G1, RMW10, 200 MeV, SOBP 10 cm



G2-250 MeV, 10 cm SOBP




G2-160 MeV, 4 cm SOBP




Results-Profile


G1, 250 MeV, SOBP 10cm, d=23.5 cm, Cross Plane




G2, RMW 91, 250 MeV, d= 23.3 cm, Cross Plane



G2, RMW 91, 250 MeV, d=23.3 cm, In Plane




G2, RMW 76, 160 MeV, SOBP 4 cm, d=6 cm, Cross Plane



Bone-Water Interface Profiles

G1, 250 MeV, d=17.9 cm



Summary

- Good agreement was found between the distal portion of the measured and calculated spread out Bragg Peaks (SOBP)
- The agreement in the proximal portion is not as good as the distal portion
- May be due to simplified assumption of single step illumination by the proton beam instead of multiple steps of RMW
- This issue will be addressed in newer versions of Eclipse.

Summary

- Agreements with the profile measurements are good within the edges of the beam
- Rapid dose fall-off beyond 20% dose level
- May be due to inadequate modeling of angular confusion of protons
- Eclipse calculated profiles in a bone phantom-water heterogeneous medium agrees well with the measurement

Conclusion

- Eclipse input data can be acquired within a reasonable time period
- Pencil-beam model of Eclipse calculates the dose distribution for protons fairly accurately, but needs further improvement
- Work is in progress to test further the accuracy of Eclipse model in heterogeneous media

Thank you for your attention

Special thanks to **Jerimy C. Polf, Ph.D.**
for providing the Monte-Carlo results
and **Hitachi staff** for their help during
our commissioning sessions

THE UNIVERSITY OF TEXAS
MD ANDERSON
CANCER CENTER
Making Cancer History®

Angular confusion and divergence

“All beam-line elements introduce a certain angular confusion to the beam. The total effect of this confusion can be measured as the shape of the penumbra behind a half-blocked beam.

Angular Confusion describes the range of directions (about some mean angle) which protons passing through any given point may have.

Angular divergence measures the variation in that mean angle over a plane perpendicular to the beam's central axis.”

From Eclipse Planning Reference Guide for Proton Algorithm

- Only angular divergence leads to sharp and magnified image
- Only Angular confusion leads to fuzzy unmagnified image