

Scintillating Plate, Block Bar, and CCD Camera as a Tool of QC Verification System in Proton Therapy at PMRC

Kiyoshi Yasuoka

*Proton Medical Research Center
(PMRC)*

*University of Tsukuba, Japan
e-mail: kiyoshi@pmrc.tsukuba.ac.jp*



Introduction

- **World Wide Spread of Particle Therapy Facilities using Rotational Gantries**
- **Importance of Quality Control (QC) for Any Equipments Relevant to Particle Therapy**
- **Importance of Isocenter (IC) Verification**



Isocenter (IC)

Definition from the Following Points of View:

1. Geometrical Position – Mechanical View
2. Patient Position – X-ray View
3. Exposure Position – Particle Beam View

Relevant to Lateral Position.



Importance of Isocenter (IC) Verification

- 1. Consistency in Center Positions of X-Ray Field and Particle Field**
- 2. Consistency in Centers of Particle Fields determined with Treatment Planning and Measurements**
- 3. Effects of Different Exposure Angle from Patient Positioning Angle**



Review of IC Verification System

1. MGH Method

- Three Dial Gauges contact with Reference Stainless Ball
 - Accuracy : $\pm 0.1 \sim \pm 0.2$ mm
 - Time to measure : 1.5 days (twice a year)
 - Verified Objects : Geometrical Center of Rotational Gantry
 - Reference : A stainless ball required

Detector directly contact with the reference ball

2. PMRC Method-1

- Imaging Plates (IP) contactless with Reference Stainless Ball
 - Accuracy : $\pm 0.1 \sim \pm 0.2$ mm
 - Time to measure : 2 hours (irregular time when it is required for QC)
 - Verified Objects : Center of Particle Irradiation Field, Center of X-ray Irradiation Field, Laser Marker
 - Reference : A stainless ball required

Detectors replaced every gantry angle out of touching with the reference ball




PMRC Method-2

Irradiation Field Images produced on a Scintillating Plate (SciP) and viewed by a CCD Camera

- **Verified Objects : Center of Particle Irradiation Field, Center of X-ray Irradiation Field, Laser Marker**
- **No Reference Ball Required**
- **No Contacts with Detectors**
- **Time to measure : Half an hour**



Verification Method



Irradiation Field Images produced on a SciP and viewed by a CCD Camera, which is setup 5-m far from IC.




Equipment

- **Plastic Scintillating Plate (SciP)**
 - NE102A
 - 200 mm x 200 mm x 1 mm^t
- **CCD Camera**
 - Hamamatsu ORCA-ER-1394
 - Nikon 105mm zoom lens
 - Hamamatsu CCD Camera Controller



Timing Chart





- 1) Beam spill width: 300ms maximum
- 2) CCD exposure-time Gate: 400ms



Analysis


Proton Irradiation Fields Detected with SciP and CCD Camera



1. **Setting up Region of Interest (ROI)**
 - 1000 pixel x 20 pixel
 - (250 mm x 5 mm)
2. **An error function fitting the edges of penumbra in each ROI.**
The edges X1 and X2 provide each center: $X_C = (X1 + X2)/2$.
3. **Scanning ROI along a beam direction every step by 20 pixels**



Results




1. A center line of proton irradiation fields is provided by a straight line fitting centers of penumbra edges obtained along beam direction at each gantry angle.
2. Accuracy in measurement of the IC position is ± 0.1 mm, which is estimated from the fitting errors for the central line of the proton irradiation fields. The accuracy at +45 degrees is ± 1.8 mm.



Comparison with Other Methods

PMRC Method 1



- In comparison with the center line of proton irradiation field at -90° , the center line at $+90^\circ$ shifts downward by 0.6 mm. This result is consistent with those obtained from PMRC Method 1 and MGH Method.



Importance of Range Verification

1. Compensators provide various shape of proton irradiation fields: small target volume and/or complicated target volume.
2. Those make it difficult to evaluate precise dose distribution in treatment planning programs even if they use a pencil beam algorithm including no scattering effects produced at steep edges of apertures and compensators.



$$R = \int_{E_0}^0 \frac{dE}{\left(\frac{dE}{dx}\right)}$$

Relevant to Longitudinal (Depth) Position



Review of Range Verification System


- Si detector scanned three-dimensionally in a water phantom
- 10 degrees-inclined imaging plate (IP) sandwiched between two triangle-shaped acrylic blocks



Range Verification using SciBar and SciP


1. SciBar&CCD Camera

- 3D-Dose Distribution in a Scintillating Block Bar → 2D Projection along Depth Direction → 1D Longitudinal Profile → Range



2. SciP&CCD Camera


- 2D-Lateral x 1D-Depth Scanning → 1D Longitudinal Profile → Range



Range

Scintillating Bar & CCD Camera


**200 MeV
No Range Shifts
No SOBP Filter**



2-D Images x 10 Slices


Scintillating Plate & CCD Camera

-10 0 +10



Range

Scintillating Plate & CCD Camera



155 MeV
No Range Shifts
No SOBP Filter

Quenching effects on
scintillating responses
provide no modification
on ranges.

M2: Beam Monitor

M3: Thimble Ionization Chamber




Quenching Effects at Higher LET Regions

- Light emission from plastic scintillating plates is reduced at higher LET regions
- Light yield reduction is 15% at Bragg peak



Range

Scintillating Plate & CCD Camera



155 MeV
No Range Shifts
50-mm SOBP Filter

Quenching effects on
scintillating responses
provide modification
on shapes.

M2: Beam Monitor

M3: Thimble Ionization Chamber



Summary

- **In PMRC Method 2**, the accuracy in measurement of the center of proton irradiation fields is the same level as MGH Method and PMRC Method 1: ± 0.1 mm.
- No reference ball and no contacts to detectors are required **in PMRC Method 2**. It is feasible to measure the center of irradiation fields for both protons and x-rays with the same detector setup and a practical short time.
- A system of scintillating plate, block bar, and CCD camera is very useful in verification of IC and ranges.

