

# Pencil Beam Scanning at PTC-H

Workshop PTCOG 45, October 7-11, Houston, Texas

Martin Bues, PhD

# Outline

- Overview pencil beam scanning
- Spot scanning with the Hitachi Probeat system
  - Performance parameters for spot scanning at PTC-H
  - Synchrotron beam extraction for pencil beam scanning
  - Scanning nozzle design
  - Dose monitoring system

# Outline (cont.)

- Beam measurements on the PTC-H scanning nozzle
  - Ion chamber Measurements of single pencil beams
  - Film measurements
- Development of treatment planning software for the PTC-H scanning nozzle
  - Commissioning of the Varian eclipse treatment planning system for proton spot scanning


# Acknowledgements

- MD Anderson
  - George Ciangaru
  - Razvan Gaza
  - Alfred Smith
- Hitachi
  - Techie Akiyama
  - Toshie Sasaki
  - Koji Matsuda
- Varian
  - Barbara Schaffner
  - Christine Ritzmann
  - Evangelos Matsinos


# Overview of Pencil Beam Scanning



# Physics of the Passive Scattering Mode of Proton Beam Delivery



# The Pencil Beam Scanning Mode of Proton Beam Delivery



# Motivation for Pencil Beam Scanning

- Fewer neutrons
- Sparing of healthy tissues proximal to the target
- Intensity modulated proton therapy (IMPT)



# Drawbacks of Pencil Beam Scanning

- More “complicated” due to large number of parameters required to specify a treatment field
- Possibility of “Interplay Effect” in the presence of organ motion

# Discrete Spot Scanning with the Hitachi ProBeat System

# Proton Therapy Center - Houston


## PTC-H

3 Rotating Gantries  
1 Fixed Port  
1 Eye Port  
1 Experimental Port

**Pencil Beam Scanning Port**

**Passive Scattering Ports**

**Experimental Port**



**Accelerator System**


**Large Field Fixed**

**Eye Port**

# Scanning Scheme

- ◆ Repeating many static irradiations
- ◆ Speedy beam switching with RF Driven Extraction technique

Dynamic scanning is achieved  
with *discrete spot scanning method*.



# Status of Pencil Beam Scanning at PTC-H

- Contract signed, May 2003
- White paper on pencil beam scanning finalized, November 2004
- Pencil beam scanning (PBS) specifications finalized, February 2006
- First beam in gantry 3, July 2006
- Turnover of PBS nozzle to PTC-H, November 2006
- First PBS patient, summer 2007




# Basic Design Parameters for PBS at PTC-H

- Triple redundancy against catastrophic failure
- Step and shoot delivery
- Maximum range: 30 cm
- Minimum range: 4 cm
- Field size: 30 x 30 cm
- Source-axis-distance (SAD): 250 cm

# Single Pencil Beam Parameters

- Spots size in air, at isocenter:
  - Range 4 cm: 11 mm
  - Range 10 cm: 6.5 mm
  - Range 20 cm: 5 mm
  - Range 30 cm: 4.5 mm

# Spot Size in Air and Water



# Single Pencil Beam Parameters (cont.)

- Difference of spot size in x and y:
  - Less than 1 mm
- Dependence of spot size on Gantry angle:
  - Less than 1mm
- Skewness:
  - Less than 0.001

# Skewness

- Measure of deviation of single spot beam from a Gaussian
- Moments of a probability function  $P(x)$ :


$$\mu_n = \int (x - \mu)^n P(x) dx$$

- **Skewness:**  $\gamma_1 = \frac{\mu_3}{(\mu_2)^{3/2}}$



# Example of Skewness

- Skewness = 0.018




# Synchrotron




Martin Bues, PhD, PTCOG 45, Houston, Pencil Beam Scanning Workshop

# Timing Parameters for PBS




# Variability of Timing Parameters for PBS



Parameters	Time Elements	Set Value Variable	Default Settings	User-configurable Settings
$\tau_{CYC}$	Repetition cycle of the magnetic field strength pattern	Yes	$\tau_{CYC} = (\tau_{FT} + \tau_{e1} + \tau_{e2})$	$(\tau_{FT} + \tau_{e1} + \tau_{e2}) \leq \tau_{CYC}$
$\tau_{FT}$	Flat top period	Yes	$\tau_{FT} = \tau_{SP}$	$\tau_{SP} \leq \tau_{FT} \leq 5\text{sec.}$
$\tau_{SP}$	Respiration synchronized Irradiation time	Yes	0.5sec.	Selectable from five preset patterns. $0.5\text{sec.} \leq \tau_{SP} \leq 5\text{sec.}$ (Only applicable for the respiration synchronized irradiation)
$\tau_{e1}$	Acceleration time	No	(Determined automatically according to the beam energy)	
$\tau_{e2}$	Deceleration time	No	(Determined automatically according to the beam energy)	


**Figure 4: Relationship between Magnetic Field Strength Pattern and Irradiation Time**

# Scanning Nozzle Design






# Scanning Nozzle Design (Cont.)



# Scanning Nozzle Design (Cont.)



# Scanning Nozzle Design (Cont.)



# Dose Monitoring Equipment

- Profile monitor
  - Multi-wired ionization chamber
  - Monitors correct incidence and shape of beam at entrance of the nozzle
- Sub dose monitor
  - Independent check for Main dose Monitor
- Main dose monitor
  - Parallel plate ionization chamber
  - Monitors number of protons delivered spot by spot

# Dose Monitoring Equipment (Cont.)

- Spot position monitor
  - Multi wired ionization chamber
  - Monitors spot position and spot size
- Spot monitor unit constraints
  - Maximum MU: 0.04 MU
  - Minimum MU: 0.005 MU
  - Resolution: 0.0001 MU = 2 pC monitor chamber charge (= 13 fC proton charge @ 175 MeV)



# Multipainting

- Locations in the target are visited by pencil beams of the same energy and deflection, **several times**
- Advantages: Reduction of dose error, Compensation for organ motion

# Example

- Uniform Irradiation of 10x10x10 cm of water, range 20 cm, to 1 Gy
  - Requires 26 Energies: 170 ... 120 MeV
  - 3.9 ... 0.1 nC proton charge for a total of 13 nC proton charge
  - ~400 spots per energy layer
  - 10400 spots in total
    - For 170 MeV: 10 pC proton charge per spot
    - i.e. 1.5 nC monitor chamber charge
    - i.e. 0.075 MU per spot
    - 30 MU for first energy layer

# Example (Cont.)

- ...
  - For 120 MeV: 0.25 pC per spot
  - i.e. 50 pC monitor chamber charge
  - i.e. 0.0025 MU per spot
  - 1 MU for last energy layer
- 
- Irradiation of 10x10x10 cube of water at a depth of 20 cm requires 100 MU

# Multipainting


- Assume Multipainting numbers of:  
8, 4, 3, 2, 2, 2, 2, 1, ...,1
- 16800 spots

# Beam Measurements

# First Measurements


- Depth dose measurements with PTW Bragg peak chamber
- Depth dose measurements with PTW Advanced Markus chamber
- Film measurements

Proton Dose Distribution ( $\text{MeV g}^{-1} \text{p}^{-1}$ ) in WATER  
 $E_p = 200 \text{ MeV}$ ,  $\sigma_p = (0.75, 0.75) \text{ cm}$ ,  $\sigma_E = 0 \%$ ,  $\text{SSD} = \infty \text{ cm}$ , Axial Cut Plane  $y = 0 \text{ cm}$



Cianguaru, et al. Med. Phys.

# PTW Bragg Peak Chamber (BPC)







# PTW Advanced Markus Chamber (AMC)




# Bragg Peak Chamber Measurement




# Markus Chamber Measurement




# Measurement Setup




# BPC Measurement vs. MCNPX



# BPC Measurement vs. MCNPX (Cont.)




# Film holder



# EBT Gafchromic Film Calibration


Optical Density - Dose Calibration Curve  
for a Gafchromic EBT film (lot # 36076-002I)  
irradiated with protons






# Dose profile perpendicular to beam



Comparison of Dose Profiles at Depth = 128 mm in Water  
measured with an Ion Chamber and a Gafchromic EBT film



# Dose profile parallel to beam (work in progress)



# Film vs. Pin Point Chamber




# Mock Commissioning of Varian Eclipse Treatment Planning System (TPS)

# Minimal TPS commissioning data requirements

- 93 CAX depth dose curves, in water for 93 energies ranging in penetration from 30 cm to 4 cm
- 2 x 5 x 93 lateral scans through CAX pencil beams in air, corresponding to 2 orthogonal scans at 5 distances from the isocenter plane


# CAX depth dose curves

Central Axis Depth Dose Curves



# Cross Profiles

Cross Profiles @ isocenter







# Simple Treatment Planning Exercise

- Simple geometric TP scenarios  
10x10x10 cm target volumes in a water phantom
- Spot placement distance of 6.5 mm
- “Optimized” and “un-optimized” treatment plans

# Scanned beam



# Un-optimized vs. optimized



# Penumbra Comparison for Varian Eclipse Plans

- Penumbra for un-optimized scanning:  
11 mm
- Penumbra for optimized scanning:  
10 mm

For reference:

- Penumbra for passive scattering:
- 10 mm

# Limit for PBS penumbra

- Spot Sigma in water x 1.3
- At center of SOBP energy of  $\sim 150$  MeV
- Spot Sigma = 6.5 mm
- 9 mm Penumbra may be achievable